

ARCS Instrument Replacement Form

Fill out the following form when replacing instruments at an ARCS site:

SV-0710M_cal

System: Skyrad NIP

Site: Manus

OLD Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) x 10 ⁻⁶ or 4 PIR coefs.	Offset	Date of Change (UTC)	Time Start UTC	Time End UTC
NIP	31876E6	Skyrad Tracker	V061012.0	8.4600	na	30/Sep/07	22:25	23:01

NEW Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) x 10 ⁻⁶ or 4 PIR coefs.	Offset	Change Logger Coefficients and Version
NIP	29937E6	Skyrad Tracker	V070930.00	7.9446	na	X

Replacement completed by: Troy Culkan

NOTES:

Old Part remains on site as spare:

Old part packed for shipment to: SGP

Other old part disposition (explain):

Other comments: